



stngco.com

**CORPORATE OFFICE**

2691 Saturn Street, Brea, CA 92821  
Tel: (714) 524-0663, Fax: (714) 364-8113



**ST&G CORPORATION**

Rm. 911 43, Yangsan-ro  
Yeongdeungpo-gu, Seoul, Korea



**MDSS GmbH**

Schiffgraben 41, 30175 Hannover, Germany



**REF: 1212T**

**Titanium Tube clamp, Ti**

Weight: 79 grams

Weight Limit: 136kg / 300lbs

Torque recommendation:

Alignment adjustment set screw: 15Nm

Pylon clamp screw: 7 Nm



**1212T**

1212T IFU REV. D (07-11-23)



stngco.com

**CORPORATE OFFICE**

2691 Saturn Street, Brea, CA 92821  
Tel: (714) 524-0663, Fax: (714) 364-8113



**ST&G CORPORATION**

Rm. 911 43, Yangsan-ro  
Yeongdeungpo-gu, Seoul, Korea



**MDSS GmbH**

Schiffgraben 41, 30175 Hannover, Germany



**REF: 1212T**

**Titanium Tube clamp, Ti**

Weight: 79 grams

Weight Limit: 136kg / 300lbs

Torque recommendation:

Alignment adjustment set screw: 15Nm

Pylon clamp screw: 7 Nm



**1212T**

1212T IFU REV. D (07-11-23)



stngco.com

**CORPORATE OFFICE**

2691 Saturn Street, Brea, CA 92821  
Tel: (714) 524-0663, Fax: (714) 364-8113



**ST&G CORPORATION**

Rm. 911 43, Yangsan-ro  
Yeongdeungpo-gu, Seoul, Korea



**MDSS GmbH**

Schiffgraben 41, 30175 Hannover, Germany



**REF: 1212T**

**Titanium Tube clamp, Ti**

Weight: 79 grams

Weight Limit: 136kg / 300lbs

Torque recommendation:

Alignment adjustment set screw: 15Nm

Pylon clamp screw: 7 Nm



**1212T**

1212T IFU REV. D (07-11-23)



stngco.com


**CORPORATE OFFICE**

2691 Saturn Street, Brea, CA 92821  
Tel: (714) 524-0663, Fax: (714) 364-8113



**ST&G CORPORATION**

Rm. 911 43, Yangsan-ro  
Yeongdeungpo-gu, Seoul, Korea



**MDSS GmbH**

Schiffgraben 41, 30175 Hannover, Germany



**REF: 1212T**

**Titanium Tube clamp, Ti**

Weight: 79 grams

Weight Limit: 136kg / 300lbs

Torque recommendation:

Alignment adjustment set screw: 15Nm

Pylon clamp screw: 7 Nm



**1212T**

1212T IFU REV. D (07-11-23)

**1. Application**

The 1212T, Tube Clamp is to be exclusively used for the prosthetic fitting of lower extremity amputations.

**2. Description**

The Tube Clamp is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply Loctite to all screws once the alignment and connection is finalized.

**3. Warranty**

Warranty applies only when the product is used as intended without modification, following all manufacturer’s recommendations. ST&G offers 2 years warranty.

2

**4. Declaration of Conformity**

ST&G as manufacturer declares that the 1212T, Tube Clamp, Ti conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

**\*Notice\***

**Tighten all the screws, which are side by side, using the torque wrench**



3

**1. Application**

The 1212T, Tube Clamp is to be exclusively used for the prosthetic fitting of lower extremity amputations.

**2. Description**

The Tube Clamp is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply Loctite to all screws once the alignment and connection is finalized.

**3. Warranty**

Warranty applies only when the product is used as intended without modification, following all manufacturer’s recommendations. ST&G offers 2 years warranty.

2

**4. Declaration of Conformity**

ST&G as manufacturer declares that the 1212T, Tube Clamp, Ti conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

**\*Notice\***

**Tighten all the screws, which are side by side, using the torque wrench**



3

**1. Application**

The 1212T, Tube Clamp is to be exclusively used for the prosthetic fitting of lower extremity amputations.

**2. Description**

The Tube Clamp is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply Loctite to all screws once the alignment and connection is finalized.

**3. Warranty**

Warranty applies only when the product is used as intended without modification, following all manufacturer’s recommendations. ST&G offers 2 years warranty.

2

**4. Declaration of Conformity**

ST&G as manufacturer declares that the 1212T, Tube Clamp, Ti conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

**\*Notice\***

**Tighten all the screws, which are side by side, using the torque wrench**



3

**1. Application**

The 1212T, Tube Clamp is to be exclusively used for the prosthetic fitting of lower extremity amputations.

**2. Description**

The Tube Clamp is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply Loctite to all screws once the alignment and connection is finalized.

**3. Warranty**

Warranty applies only when the product is used as intended without modification, following all manufacturer’s recommendations. ST&G offers 2 years warranty.

2

**4. Declaration of Conformity**

ST&G as manufacturer declares that the 1212T, Tube Clamp, Ti conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty, or liability. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

**\*Notice\***

**Tighten all the screws, which are side by side, using the torque wrench**



3